

PHILIPPINE HEART CENTER
East Avenue, Quezon City

Minutes of the Regular BAC Meeting and Pre Bid Conference
April 20, 2021 9:00am, 5th Amphitheatre, PHC

BAC.Meeting.016.21

1. Attendance:

The meeting was participated by members of the BAC, TWG and its Secretariat as follows:

<p>Chairman : Marietta A. Velasco, RN, MAN – on leave</p> <p>Vice-Chairman : Antonio D. Pascual, MD</p> <p>Members : Peter San Diego, Jr. MD Rex B. Garde, RMT – on leave Agnes L. Balid - absent Agnes Grande Daisy Algenio</p>	<p style="text-align: right;">TWG Members</p> <p>Chairman : Marcial Magleo – on leave</p> <p>Vice Chairman : Marissa Sarmiento - absent</p> <p>Members : Engr. Dionisio Eoy : Maria Charisse Y. Magallanes, RN : Annalyn Ollado</p> <p style="text-align: right;">Secretariat : Jeanette Z. Burillo Rachelle H. Cortez Joy Y. Lacanilao Grace Bernadette T. Hugo</p>
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

I.1. Call to Order

I.2. The meeting was called to order at 9:37am, presided by the Vice Chairman Antonio Pascual, MD.

The purpose of the Pre Bid Conference was reiterated by Dr. Pascual stating that the Pre Bid Conference serves as an avenue to clarify and verify concerns regarding such Project.

II. Agenda: Pre Bid Conference

Project Title	Clarifications/Modifications/Remarks
1 Unit Instrument Washer End-user: OR/NSO Christopher Ian Joseph M. De Guzman ABC = P5M Prospective bidders: <ul style="list-style-type: none"> - MedMaster, Inc – Joyce J. Mendoza - MEDICOTEK, Inc. – Kriska Kahulugan - AMI Equipment Services & Solutions – Allen Olidan 	<p style="text-align: center;">Corporate Fund Included in the APP for CY2021</p> <p>No prospective bidders bought bidding documents</p> <p>: Medmaster: On Technical Specifications:</p> <p>> Dimension: requested for a minimal adjustment, 2 inches bigger.</p> <p>Equipment: Width: 724mm; depth: 724mm; height: 1981mm</p> <p>Chamber: width: 635mm; depth: 642mm; height: 642mm</p> <p>> Dr. Pascual asked the End-user if the requested increase in the DIMENSION and CHAMBER size is acceptable considering the existing space.</p> <p>> Mr. De Guzman (End-user) reiterated that the area is already tight where the washer can be placed. His concern is that with the 2 inches adjustment requested by MedMaster, it may hamper a possible additional equipment in the future.</p> <p>> Medicotek and AMI: on 1.10 Door – pass thru, sliding door.</p> <p>> AMI offered a possibility of slight renovation of the</p>

	<p>area so for their machine to fit in which was attested by Mr. De Guzman that they can only allow slight renovation so not to disrupt the operation in the area.</p> <p>> 1.2 Load Capacity: MedMaster inquired on 12 basket DIN; 6 level rack.</p> <ul style="list-style-type: none"> > End-user will stick to their requirement based on the existing machine that they have. > Medmaster and AMI – requested for a 4 basket DIN. <p>> 1.4 Control Panel – Medmaster requested for a 5.7 inch touch screen, certified also that interfaces are clearly visible with the 5.7 inch touch screen.</p> <ul style="list-style-type: none"> - Polycarbonate as protective layer for the touch screen, requested to indicate <u>“or its equivalent protective layer”</u> <p>> Consumables as reiterated by Mr. De Guzman should not be limited to specific brands.</p> <p>> Dr. Pascual verified with the prospective bidders if their consumables are not specific to their unit. MedMaster stated that the issue here is that usually consumables (detergents or reagents) can be programmed to the machine.</p> <p>Medmaster suggested to indicate/include in the Terms and Condition: The machine should be compatible to any machine chemical so for them to set it to the machine.</p> <p>> End-user noted the requests and will try to consider if the requested adjustment can be accommodated. All changes in the Technical Specifications will be posted in a supplemental bid bulletin.</p> <p>Announcement:</p> <ol style="list-style-type: none"> 1. Deadline for the submission of the Eligibility Requirement for the issuance of Certificate of Simplified Supplier’s Registration (CSSR) is on April 23, 2021 until 4pm at the BAC Office. 2. Opening of Bids: May 4, 2021 at 9am.
<p>2 units Defibrillator with Pacer and Cart</p> <p>End-user: OR/NSO – Christopher Ian Joseph M. De Guzman</p> <p>ABC = P1.8M</p> <p>Prospective bidders:</p> <ul style="list-style-type: none"> - IDS Medical System Phils. – Shane Domalaon - Variance Trading Corp – Arthur Gillera - Pinnacle Medical - Tehcnomed International - - MEDEV – Kenneth Mercado 	<p style="text-align: center;">Corporate Fund</p> <p style="text-align: center;">Included in the APP for CY2021</p> <p style="text-align: center;">No prospective bidders bought bidding documents</p> <p>On Technical Specifications:</p> <p>Internal Paddles: Pediatrics, Mr. De Guzman requested for a smaller size diameter. Will amend what was stated in the original Technical Specification from 3.5 – 4.5cm in diameter to 2.5cm to 3.5cm.</p> <p>Variance: 2.2.a: Pacer pad connector compatible with multiple brands, Ms. Magallanes preferred a “universal” pads so to fit into the connector in the machine. All suppliers attested that their machine is specific to its own pad connector. This is to be verified to the ICUs with their experience on the existing machines.</p> <p>2.2.b. Two sets for pediatrics and adult, Variance has a 2</p>

in 1 pad size to be configured.

> Other suppliers has 1 size or 2 in 1 works for both adult and pediatrics.

Electrodes:

> End-user noted the requests and will try to consider if the requested adjustment can be accommodated. All changes in the Technical Specifications will be posted in a supplemental bid bulletin.

Announcement:

1. Deadline for the submission of the Eligibility Requirement for the issuance of Certificate of Simplified Supplier's Registration (CSSR) is on April 23, 2021 until 4pm at the BAC Office.

2. Opening of Bids: May 4, 2021 at 9am.

Other matters:

1. Motion for Reconsideration (MR) filed by the following Companies:

1.1. Medilines Distributors, Inc. 1 pc Mobile C-arm Fluoroscopic Unit

- To schedule Post Qualification, and ask Supplier to secure documentary evidence to serve as proof of their compliance.

1.2. Dynamed Healthcare, Inc. – 3 units High Flow Nasal Cannula Oxygen Therapy Machine

- As per inquiry with the End-user, no requirement for a product demonstration thus proceeded with the Post Qualification process. BAC reiterated the basic requirement on the Statement of Compliance, following the PASS or FAIL criteria. RESOLUTION: Motion for Reconsideration is DENIED. Will proceed to the next Lowest Calculated Bid.


2. Determination of Lowest Calculated Bidder (LCB) relative to the Project, re: Pharmaceutical Supplies.

3. Schedule of Post Qualification, re: Pharmaceutical Supplies.


- set on Friday, April 23, 2021, Ms. Agnes Grande as in-charge.

4. Meeting adjourned at 11:05am.

Prepared by:


JEANETTE Z. BURILLO, RL
Chairman, BAC Secretariat

Approved by:


ANTONIO D. PASCUAL, MD
Vice Chairman, BAC for Pharma Supplies &
Med Equipment