

PHILIPPINE HEART CENTER
East Avenue, Quezon City

Minutes of the Regular BAC Meeting and Pre-Bid Conference
February 2, 2021, 9:00am, 5th Floor, Meeting Room C, PHC

BAC Meeting:006.21)

1. Attendance:

The meeting was participated by members of the BAC, TWG and its Secretariat as follows:

Chairman : Marietta A. Velasco, RN, MAN - ACI
Vice-Chairman : Antonio D. Pascual, MD
Members : Peter San Diego, Jr. MD - ACI
Rex B. Garde, RMT
Agnes L. Balid
Agnes Grande

TWG Members
Chairman : Marcial Magleo
Vice Chairman : Marissa Sarmiento
Members : Engr Dionisio Eoy
Maria Charisse Y. Magallanes, RN
Annalyn Ollado

Secretariat : Jeanette Z. Burillo
Rachelle H. Cortez
Joy Y. Lacanilao
Grace Bernadette T. Hugo

2. BAC Members and the Technical Working Group (TWG) proceeded with the Pre-Bid Conference at exactly 9:30am facilitated by the Co-Chairman, Antonio D. Pascual, MD.

3. **Announcement:** Submission of Certificate of Simplified Suppliers Registration (CSSR) February 5, 2021 until 4pm at the BAC Office.
:Opening of Bids: February 16, 2021

II. Agenda: Pre-Bid

	Project Title	Clarifications/Modification
1	<p>3 units High Flow Nasal Cannula Oxygen Therapy Machine End-user: Pulmonary ABC = P1,200,000.00</p> <p>Prospective bidder who bought bidding documents:</p> <p>1. Biomedica Healthcare System – Eduardo Giron</p> <p>Observers: Josmeth Enterprises – Dennis Alfonso Dynamed Healthcare Inc. – Hazel Candidato Enhance Medical and Industria Enterprises – Bernard Esperon HealthTrust Enterprises – Fedelito Ramos Chemvalley - MEDEV – Wendell Madriaga Southside Biomedica – Aimee Tagoon</p>	<p>> BFAD as requirement of bidding. As per End-user, BFAD is not a requirement.</p> <p>> One bidder stated that their unit has a disinfection functionality but was considered as an additional feature since it is not stated in the Technical Specifications.</p> <p>> 3.2.1. Flowrate – 2-60 LPM, bidders has 2-80LPM</p> <p>> FiO2 21-95 (+/-5%) 25-95 – still within the required specification</p> <p>3.3.2.2. ALARM – as long as there is a sensor or alarm indicator.</p> <p>3.4.1.1. Display Screen 2.3” to 3.4” Dynamed has screen measuring 4.3. - End-user prefers their stated specification</p> <p>3.5.2 Flowmeter – highflow system, two bidders have an integrated flowmeters, can be considered by the End-user.</p>

		<p>4.1. Autovolt 110-220 volts/60 hz – bidder has 220volts, majority has autovolt. End-user prefers the indication of autovolt 110-220 volts/60hz as specified in the TS.</p> <p>* Dr Pascual advised that as long as it can fit with the device and it will not affect its functionality. Requested the End-user to discuss the size.</p> <p>* End-user also specified that what's indicated in the Technical Specification is their minimal requirement.</p> <p>* All changes will be announced via to Supplemental Bid Bulletin</p>
2	<p>1 Lot 20 Units 12 Lead ECG with ECG Management System ABC = P8M Enduser: EPS/Eden Gabriel, MD/Desmond Curameng</p> <p>Prospective bidders who bought bidding documents: Dynamed Healthcare– Franz Blaza</p> <p>Observers: Technomed RG Meditron Saviour MeDevices ICare JOSMEF Enterprises</p>	<p>3.1.G – Arm Lead reversal – visual indicator is a requirement.</p> <p>3.2.A – compatibility to the existing units and interface with the PHC-HIS and the ECG Management System.</p> <p>- One bidder provides their own software rather than the Philips Page Writer TC20 as stated in the TS. But assures that the integration can be facilitated.</p> <p>- PHC uses the DICOM connectivity, thus TS has to state "DICOM ready" rather than stating proprietary connectivity. This concern has to be discussed with the IT</p> <p>3.2.L – Manual or automatic readings</p> <p>3.2.Q – XML and TIFF not included, PDF is sufficient.</p> <p>8.4 Calibration of Machine as inclusion to the warranty of preventive maintenance for 3 years</p> <p>> Delivery Period: 90 days</p> <p>* End-user also specified that what's indicated in the TS is their minimal requirement.</p> <p>* All changes will be subject to Supplemental Bid Bulletin</p>
3	<p>1 pc Mobile C-Arm Fluoroscopic Unit ABC= 9M End-user: CV Radiological Sciences/ Dr. Rosalinda Estrella/Ramon Sacluti</p> <p>Observers: Medilines – Mary Grace Hernandez Health Solutions – Princess Joy Fernandez Siemens Healthcare- Cecilyn Failona Josmeth Medical – Carla Bautista Variance Trading – Jovilyn Bautista GE Healthcare – Sheryl Mesa Technomed – Mariola Ramirez</p>	<p>> IEC – to be corrected by the End-user – 60601 and 60613</p> <p>> Heat Storage capacity of at least 1,100,000 HU</p> <p>B.4. Swivel Range at least 12⁰ – to range $\pm 10^0$</p> <p>B.7. Depth of the C-arm of 73cm – to consider at least 66cm</p> <p>B.8. Flat detector distance – to consider at least from 66cm</p> <p>B.9. FD Distance 100cm – to consider at least 99.3cm</p> <p>D.1. 2.3 W to 2.3 KW</p>

	Juniper - Rizalyn	<p>Other requirements:</p> <ul style="list-style-type: none"> - Autovolt, Environmental Factors, Disposable consumables and Warranty to be discuss by the End-user and the EMD > Delivery Period: 90 calendar days > Bidder ask to provide communication on their considerations so to be discuss by the End-user <p>* All changes will be subject to Supplemental Bid Bulletin</p>
4	<p>1 unit Video Conferencing System (Virtual OR Mobile Machine)</p> <p>ABC = 8M</p> <p>End-user: OR – Edgardo Mendoza HIS – Peter San Diego, MD/ Renato Doctor</p> <p>Observers: MIS – Joan Lopez MedEqual – Mark Canlas PredICT – Romwell San Dela Torre BGP Solutions – Rodel Asa</p>	<p>3.A. As per the HIS, they preferred with the embedded IP platform.</p> <p>3.B. To include option to stream online</p> <p>5.1 Environmental Factors</p> <p>Temperature – 0-50 degrees</p> <p>Humidity – 10 – 93% can be considered</p> <p>> Adapter: HDMI and to provide with adapter for DDI</p> <p>> Proof of Concept (demo) – supplier to bid and make sure if they are qualified</p> <p>> Delivery Period: 90 days</p> <p>> SLCC – Similar projects and medical equipment can be considered</p> <p>* All changes will be subject to Supplemental Bid Bulletin</p>

IV. Other matters:

1. 2nd LCB – Post qualification process done and they have provided needed documentations and is compliant.
2. Update on Medilines: no more MR forwarded after our communication.
3. UNILAB communication, sole bidder. Cannot comply delivery of Furosimide. Stated their willingness to pay for the penalty imposed until such time rebid is completed.
4. Meeting was adjourned: 11:45am

Prepared by:



JEANETTE Z. BURILLO, RL
Chairman, BAC Secretariat

Approved by:



ANTONIO D. PASCUAL, MD
Co-Chairman, BAC for Pharma Supplies & Med Equipment