



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 002-2023

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 3 units Intra-Aortic Balloon Pump - ITB No. 004.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued schedule of the Opening of Bids as stipulated in the bidding documents, thus, revisions/changes are introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 41	1.7. Pneumatics 1.7.1. Stepper motor/scroll compressor driven bellows using UPS-graded helium	1.7. Pneumatics 1.7.1. Stepper motor/scroll compressor driven bellows using USP-graded helium
Page 42	5. Standards, Safety and Training 5.3. Regulations" DOH Philippines	5. Standards, Safety and Training 5.3. Valid Certificate of Medical Device Registration from FDA Philippines
	6. Documentation 6.2. Certificate of calibration ad inspection	6. Documentation 6.2. Certificate of calibration ad inspection from the Principal/Manufacturer

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 17th day of February 2023, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.

Vice-Chairman, BAC for Pharmaceutical Supplies and Med Equipment