

## PHILIPPINE HEART CENTER East Avenue, Quezon City

## SUPPLEMENTAL BID BULLETIN No. 003-2023

SUBJECT

**Revisions in the Bidding Documents** 

PROJECT/S

1 unit Ultra Low Freezer - ITB No. 002.23

\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued schedule of the Opening of Bids as stipulated in the bidding documents, thus, revisions/changes are introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 39	1- Technical Specifications:  1.3. System should have programmable operating temperature range (ambient temperature/from - 50°C to -86°C with 1°C increment  1.12. Freezer must use CFC/HCFC free refrigerant	1- Technical Specifications: 1.3. System should have programmable operating temperature range -40°C to -50°C to -85 °C to -87°C  1.12. Non toxic refrigerant
Page 41	Documentation     1.5. Certificate of calibration of the machine and tools for calibration	Documentation     1.5. Certificate of calibration of the machine from the Manufacturer and tools for calibration from the trained Engineers from Manufacturer of the Supplier.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 17th day of February 2023, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.

Vice-Chairman, BAC for Pharmaceutical Supplies and Med Equipment