

PHILIPPINE HEART CENTER East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 004-2022

SUBJECT: Revisions in the Bidding Documents

PROJECT/S: 5 units Transport Monitor (Cardiac Monitor) - ITB No. 105.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally preissued bidding documents as mentioned above, revisions are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 10	Delivery Period- 30 calendar days	Delivery Period- 60 calendar days
Page 37	Technical Specifications: 1.3. Minimum of 4-12 numeric data and/or waveform display, configurable screen display, touch screen	1.3. At least 8 numeric data and/or waveform display, configurable screen display, touch screen
	1.8. Built-in printer	1.8. Built-in printer: 3 channels Paper: Thermal
Page 38	1.11. Non Pulse Oximetry (SPO2) Invasive Blood Pressure 1.11.3. Display Range: 10-200 mmHg	1.11.3. Display Range: Minimum: 10 mmHg Maximum: 270-300 mmHg
	5. Standard, Safety and Training 5.5. Minimum of warranty of 5 years on parts and services	5.5. Minimum of warranty of 3 years on parts and services 5.5.1. Preventive maintenance every quarter for 3 years
	6. Documentation 6.4. List of Equipments available for providing calibration and routine maintenance support as per manufacturer documentation in service/technical manual	6.4. List of Equipment . available for providing calibration and routine maintenance support as per manufacturer documentation in service/technical manual
Page 41	Checklist of Technical & Financial Documents (1) . Conformity with the Technical Specifications include the following: • Certificate of Production/Delivery Schedule • Certificate of After Sales/Parts	 Certificate of Delivery Schedule of the Product Certificate of After Sales Services

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 4th day of January 2022, PHC.

MARIETTA A. VELASCO, RN, MAN.

Chairman, BAC for Pharmaceutical Supplies & Med Equipment