



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 015-2021**

**SUBJECT** : Revisions in the Bidding Documents  
**PROJECT/S** : **2 units Defibrillator with Pacer and Cart -**  
**ITB No. 017.21**


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	<b>Technical Specification</b> 1.9 Internal paddle electrodes (with 2 internal paddles each unit: 1 adult and 1 pediatric set paddles for each unit) a. Standard sizes for adult, pediatric, neonatal Size: Adult: 6 - 7.5cm in diameter / pediatrics: 3.5 - 4.5cm in diameter	1.9 Internal paddle electrodes (with 2 internal paddles each unit: 1 adult and 1 pediatric set paddles for each unit) a. Standard sizes for adult, pediatric, neonatal Size: Adult: 6 - 7.5cm in diameter / pediatrics: 2.5 - 3.5cm in diameter
Page 38	<b>System Configuration Accessories, spares and consumables</b> 2.1 System as specified 2.2 Consumables a. Pacer pad connector compatible with multiple brands b. with 2 sets of pacer pads for adult and pediatrics c. electrodes set of 20 pcs to be included upon delivery	2.1 System as specified 2.2 Consumables a. Pacer pad connector b. with 4 sets of pacer pads c. electrodes set of 20 pcs. to be included upon delivery

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 27<sup>th</sup> day of April 2021, BAC Office, PHC.

  
**MARIETTA A. VELASCO, RN, MAN**  
Chairman, BAC for Pharma Supplies and Med Equipment