



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 016-2023**

**SUBJECT : Revised/Amended Technical Specifications**

**PROJECT/S : 9 units Stretcher (Negotiated Procurement-  
2 Failed Biddings) - ITB No. 007.23**

\*\*\*\*\*  
To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are introduced.

<b>Description of Function</b>	
Stretcher is used for transfer or transport of patient for procedure and diagnostic test.	
<b>Operational Requirements</b>	
System complete with all accessories as specified in the Technical Specifications below.	
<b>Technical Specifications</b>	
1	Hydraulic height adjustment
2	Four (4) wheel casters with central locking system or its equivalent.
	2.1. Wheel diameter: 7"- 8"
	2.2. Non marking rubber is used for wheels to prevent casters from leaving marks on floor
3	Dual foot pedal control for Trendelenburg and reverse
	3.1. lower range: 0
	3.2. higher range: 16 - 20 °
4	Stretcher body: made of stainless steel or powder coated materials
5	Weight capacity: safe working load 300 kg
6	Height adjustment:
	6.1. High position: maximum 100 cm
	6.2. Lowest position: maximum 54 cm
7	Width: 75 cm – 85 cm
8	Length: 200 cm – 220 cm (bumpers are included in the measurement)
9	Side rails: fixable and collapsible with double locking system for patient's safety or its equivalent
10	Power assist fowler back rest

11	Mattress:
	11.1. washable and waterproof cover to keep mattress from fluid penetration
	11.2. thickness: 3 – 4 inches
12	With rubber or polymer plastic bumpers:
13	Must have the following accessories
	13.1. a. Safety belt: Push button buckle belt type (body and lower extremities)
	b. Strap type: Coated polyester vinyl, fluid resistant and can be cleaned and sanitized
	c. Strap width: 3 “- 4”
	13.2 Detachable tray with strap to hold monitoring devices weighing at least 5Kg.
	13.3. Compartment for patient's belongings
	13.4. Control brake
	13.5. Integrated portable oxygen holder oblique (stainless)
	13.6. Adjustable IV pole *height) with 2-4 hooks, removable or collapsible
	13.7 Urine bag holder
	13.8. Transport handle head and lower part
14	With 2 years warranty for parts and services
15	Preventive maintenance twice a year for 3 years
16	ISO 13485
17	Provision of user's manual both hard and soft copy upon delivery
18	60 days schedule of delivery
19	Duly Notarized Certificate of Exclusive Distributorship
20	Product demonstration of the same unit during post qualification.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 10<sup>th</sup> day of March 2023, BAC Office, PHC.

**MARIETTA A. VELASCO, RN, MAN, MBAH**  
Chairman, BAC for Pharmaceutical Supplies & Med Equipment