



PHILIPPINE HEART
PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 016-2024

SUBJECT : Revised Technical Specifications of the Project

PROJECT: 1 Lot Cardiac Monitoring System with Central Monitor, Defibrillator with Pacer and Cart, Transport Monitor - ITB No. 016.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, documents the originally pre-issued bidding documents as mentioned above, revision/amendment is hereby introduced.

Page No/ Item No.	PHC- Technical Specifications	Clarifications/Amended "Should be read as"
Bid Data Sheet Sec III 5.3	For this purpose, contracts similar to the Project shall be: a. Cardiac Monitor with either of the following: 1. Central Monitoring 2. Transport Monitoring 3. Defibrillator b. Completed within the past 3 years prior to the deadline for submission and receipt of bids.	No Change
1.12.2.	Each bedside monitor must have UPS that can sustain operation for at least 60 minutes	Each bedside monitor must have AVR 2.0 KVA As per the recommendation of the PHC Engineering Division /BIOMED
3.2	Data storage for at least 48 hours	No Change
3.9	Output energy: 2 to 270 to 360 joules ; adjustable setting	No Change
3.12.1.	Quick recovery ECG baseline at less than 3 seconds	No Change

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 20th day of February 2024, BAC Office, PHC.

ANTONIO D. PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment