



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 017-2020

SUBJECT : Revised Bidding Documents

PROJECT : 1 Lot Cardiac Monitoring Telemetry System - ITB No. 027.20

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre- issued bidding documents as mentioned above, revisions are hereby introduced:

| Item No. | PHC – Technical Specifications FROM | Clarifications/Modifications (Should be read as) TO |
|----------|---|---|
| Page 75 | iii. NIBP | iii. NIBP |
| | 8. With variable cuff sizes fit for adult patients | 8. With variable cuff sizes fit for adult patients |
| | a. 12 cm x 22 cm (small adult with arm circumference 22-26cm) – 15 pcs. | a. 12 cm x 22 cm (small adult with arm circumference 18-26cm) – 15 pcs |
| | b. 16 cm x 30 cm (regular adult with arm circumference 27 – 34 cm) – 70 pcs | b. 16 cm x 30 cm (regular adult with arm circumference 23 – 35cm) – 70 pcs |
| | c. 16 cm x 36 cm (large adult with arm circumference 35 – 42 cm) 30 pcs. | c. 16 cm x 36 cm (large adult with arm circumference 33 – 47cm) – 30 pcs |
| | d. 16 cm x 42 cm (extra-large adult with arm circumference 45 – 52 cm) 15 pcs. | d. 16 cm x 42 cm (extra-large adult with arm circumference 45 -66cm) – 15 pcs |
| | c. Battery | c. Battery |
| | 1. Rechargeable Dual Cell Lithium-Ion battery | 1. Rechargeable Lithium-Ion battery or NIMH or Alkaline |
| Page 76 | III. Central Monitoring Stations | III. Central Monitoring Stations |
| | 1. Master/Slave structure with the following set up | 1. Master/Slave structure with LCD display monitors of at least 24”in size and resolution of at least 1280 x 1024 |
| | 2. Up to four (4) LCD display monitors of at least 24” in size and resolution of at least 1280 x 1024 per set of central monitoring station | delete |
| | V. Power Supply | V. Power Supply |
| | 3. Voltage corrector/stabilizer of appropriate ratings ISI specifications (Input: 160-260V; Output 220-240V and 60Hz) | (combined # 3 and 4) with suitable UPS with Voltage stabilizer of appropriate ratings ISI specifications (Input: 160-260V; Output 220-240V and 60Hz) and with maintenance-free batteries for minimum one (1) hour back-up |
| | 4. Supplied with suitable UPS with maintenance-free batteries for minimum one (1) hour back-up | |

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 29th day of September 2020, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN
Chairman, BAC for Pharma Supplies and Med Equipment