

SUPPLEMENTAL BID BULLETIN No. 019-2021

SUBJECT: Revisions in the Bidding Documents

PROJECT/S: 38 units Wheel Chairs - ITB No. 022.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally preissued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	Technical Specification 1. Width of seat = 40 cm - 560 cm	1. Width of seat = $60 \text{ cm} - 70 \text{ cm}$
Page 38	2. Weight capacity = 120 - 150 kg	2. Weight capacity = 120 – 250 kg
	7. Provision of the following: 7.2. IV Pole 7.3. Oxygen portable holder 7.4. Patient's strap (buckle type)	 7. Provision of the following: 7.2. IV Pole adjustable- cross type (4 hooks) 7.3. Oxygen portable holder = O2 portable Size D 7.4. Patient's strap (buckle type) = adjustable patient's strap
		Additional Specifications: 1. Convertible to bed
		2. Wheels with brake
		3. Body materials: stainless
		4. Brochure mustbe available
		5. Warranty ≡ 3 years on parts and service
		6. Preventive Maintenance = every 6 mos to 2 years
		7. Availability of spare parts/accessories 5-7 years in the market
		8. Certification from International Safety Standards
		9. Delivery Period = 30 calendar days

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this ^{2nd} day of June 2021, BAC Office, PHC.

MARIETTA AVELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment