



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 021-2020

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : **3 units Defibrillator with External Pacer and Cart - ITB No. 034.20**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre- issued bidding documents as mentioned above, revisions are hereby introduced:

Item No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
35.a	Capnography: Warm up time at least less than 10 seconds; capable for intubated or non-intubated patients (accessories provided)	Capnography: Warm up time of 10 seconds or less; capable for intubated or non-intubated patients (accessories/cables provided)
39.a	Size: Adult: 7.5 cm in diameter/pediatrics: 3.5 cm in diameter	Size: Adult: 7.5 cm in diameter (+/- 1.0 cm)/ Pediatrics: 3.5 cm in diameter (+/-1.0 cm)
39.b	Handle with discharge buttons	Handle with discharge buttons (built in/integrated)
3.18	Built-in handle that can be mounted on bed or stretcher railing	Built-in handle, hook, or clamp that can be used to mount or secure on bed or stretcher railing

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 3rd day of November 2020, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN
chairman, BAC for Pharma Supplies and Med Equipment