



PHILIPPINE HEART CENTER

East Avenue, Quezon City

## SUPPLEMENTAL BID BULLETIN No. 022-2021

**SUBJECT** : Revisions in the Bidding Documents

**PROJECT/S** : **2 units Sternal Saw (Battery-Operated) - ITB No. 027.21**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre- issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Modifications/Amendment (Should be read as) TO
37	<b>Technical Specifications</b>	<b>Technical Specifications</b>
	1. Oscillating Saw	1. Oscillating Saw
	d. Stainless Steel and Titanium construction material	d. Stainless Steel <b>and/or</b> Titanium construction material
	2. 4 pieces Sternum Blades per unit	2. 4 pieces Sternum Blades per unit
	a. 2 (pcs.) Heavy duty sternum blade (Saw Blade) 35mm/75 mm/0.7mm/0.9mm	a. 2 (pcs.) Heavy duty sternum blade (Saw Blade) <b>35mm/75 to 82mm/0.51 to 0.7mm/0.9 mm</b>
	b. 1 (pc.) Heavy duty sternum blade (Saw Blade) 50mm/25mm/ 0.7mm/0.9 mm	b. 1 (pc.) Heavy duty sternum blade (Saw Blade) <b>50mm/23 to 25mm/0.51 to 0.7mm/0.9 mm</b>
38	3. Rechargeable high capacity lithium Ion on Nimh Battery	3. Rechargeable high capacity lithium Ion or Nimh Battery
	4. Sterile Battery Funnel for aseptic loading of battery into hand piece	4. Sterile Battery Funnel for aseptic loading of battery into hand piece, <b>if applicable</b>
	5. Battery Removal Device	5. Battery Removal Device, <b>if applicable</b>
	6. Instrument Container per unit	6. <b>Rigid instrument container system with integrated permanent germ barrier made of PTFE designed and validated for 5,000 sterilization cycles (1 system per unit)</b>
	a. Dimension: 592 x 274 x 135 mm (minimum)	a. Dimension: <b>275 to 592mm x 250 to 274mm x 125 to 135mm</b>
	b. Sterilization stainless steel container plus insert tray	b. Sterilizable stainless steel container plus insert tray
	- Perforated basket: 485 z 235 x 106 mm (minimum)	- Perforated basket: <b>244 to 485mm x 234 to 253mm x 105 to 106mm</b>
	- Perforated basket lid: 489 z 257 mm (minimum)	- Perforated basket lid: <b>244 to 489mm x 234 to 257mm</b>
	c. Fixation devices	c. With Fixation Devices <b>inside the basket</b>
		<b>7. Other Requirements:</b>
		<b>a. Warranty: Three (3) years in parts and services upon delivery</b>
		<b>b. Availability of Spare Parts: at least 5 years</b>
		<b>c. Preventive maintenance every 6 months for 3 years</b>




Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
		d. Available loaner/back up unit in case of unit malfunction/servicing
	Delivery: 30 calendar days	e. Delivery: 60 calendar days
		f. Duly Notarized Certificate of Exclusive and/or Authorized Distributorship
		g. With FDA Certifications
		h. Training for the End-users and PHC BIOMED
		i. With safety certifications of the product

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 14<sup>th</sup> day of June 2021, BAC Office.



**MARIETTA A. VELASCO, RN, MAN.**  
Chairman, BAC for Pharma Supplies and Med. Equipment

