

## PHILIPPINE HEART CENTER East Avenue, Quezon City

## SUPPLEMENTAL BID BULLETIN No. 022-2023

**SUBJECT** 

: Revisions in the Technical Specifications

PROJECT/S: 1 unit Ultra-Low Freezer (Re-Bid) - ITB No. 027.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 39	1. Technical Specifications: 1.2. Exterior: Width: 35-40 inches Depth: 30-40 inches	<ol> <li>Technical Specifications:</li> <li>Exterior: Width: 35-45 inches</li> <li>Depth: 30-42 inches</li> </ol>
	<ul><li>1.4. (Ambient Temperature from 10°C to 32°C</li><li>1.9. With 4-5 individual inner door to open independently</li></ul>	<ul><li>1.4. (Ambient Temperature from 15°C to 32°C</li><li>1.9. With 2-4 inner door to open independently</li></ul>
Page 40 / 2 & 4	2. Environmental Factors: 2.2. The unit shall be capable of operating in	2. Environmental Factors: 2.2. The unit shall be capable of operating in ambient temperature of 15-32°C and relative humidity of less than 70%.
	Standards, Safety and Training     4.7. Availability of equipment for post-qualification within 3 days from receipt of notice.	4. Standards, Safety and Training 4.7. Availability of equipment for post-qualification within 5 days from receipt of notice.
	4.13. Must provide technical assistance that is accessible 27/7 including weekends and holidays and availability of technical assistance to respond within 2 hours after call.	accessible 24/7 including weekends and holidays

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 2<sup>nd</sup> day of May 2023, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH

Chairman, BAC for Pharmaceutical Supplies & Med Equipment