



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 024-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Intra Aortic Balloon Pump - ITB No. 025.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre- issued bidding documents as mentioned above, revisions are hereby introduced:

| Item No./ Page No. | PHC – Technical Specifications FROM | Modifications/Amendment (Should be read as) TO |
|---------------------------|--|--|
| 38 | 1.f.2 Size: at least 13 inch diagonal display | f.2 Size: at 12-13 inch diagonal display |
| 38 | 1.j. Helium Cylinder can be refilled without stopping pump operation of machine | j. Helium Cylinder can be refilled/replaced without stopping pump operation of machine |
| 39 | 4.f Timing reference display: displayed in numeral timing settings in both operating modes as well as bar graphs displaying inflate/deflate events in operator mode. | 4.f Timing reference display: displayed in numeral timing settings in both operating modes as well as bar graphs/ inflation deflation intervals displaying inflate/deflate events in operator mode. |
| 39 | 7.a Stepper motor driven bellows using UPS-grade helium | 7.a Stepper motor/ Scroll compressor driven bellows using UPS-grade helium |
| 39 | 7.b Pumping volume: 0.5cc to 50 cc, adjustable in 0.5cc increments | 7.b Pumping volume: 0.5cc to 50 cc, adjustable in 0.5cc increments/ augmentation level 1-10 |
| 38 | 9.a.3 Weissler: ECG only, inflation timing based on systolic time intervals. | 9.a.3 ECG, inflation timing based on systolic time intervals. |
| 40 | 9.b.3 Weissler: ECG only, inflation timing based on systolic time intervals. | 9.b.3 ECG, inflation timing based on systolic time intervals. |
| | 3. Other Requirements | 3. Other Requirements |
| 40 | 3.1 Regulatory approval/certification FDA/CEUL/ BIS approved product | 3.1 Regulatory approval/certification FDA Philippines/ CEUL/BIS approved product |

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 14th day of June 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment

