



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 032-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Tissue Embedding Center - ITB No. 036.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications. (Should be read as) TO
Page 38	1 unit Treadmill Stress Test Machine	1 unit Tissue Embedding Center
	Technical Specifications- 9. Free installation and must provide automatic voltage regulator and UPS	9. Free installation and must provide automatic voltage regulator and uninterrupted power supply that can last for at least 30 mins.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 8th day of July 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment