



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 034-2022 (B)

SUBJECT : Revision in the Bidding Documents

PROJECT : 6 units Transport Monitor with Defibrillator - ITB.038.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specification as mentioned above, revisions/changes are hereby introduced.

ITB No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	1.3. Color LCD screen: at least 6.5 – 7.5 inches display; minimum resolution of 640 x 480 pixels	1.3. Color LCD screen: at least 6.5 – 8.5 inches display ; minimum resolution of 640 x 480 pixels
Page 37	1.4 Monitoring Parameters: 1.4.1.4. Simultaneous analysis of at least 4 leads (user defined)	1.4 Monitoring Parameters: 1.4.1.4. Simultaneous analysis of 3-5 leads (user defined)
Page 38	1.4.3.2.. Range: Systolic: 25 mmHg - 290 mmHg Diastolic: 10 mmHg – 210 mmHg Mean: 15 mmHg – 270 mmHg	1.4.3.2.. Range: Mean: 15 mmHg – 270 mmHg- To be deleted
Page 39	4. Power Supply 4.2. Re-settable over-current breaker shall be fitted for protection	4.2. To be deleted
	4.5 Full charge running time: 36-48 hours	4.5 Full charge running time: 4.5.1. Monitoring: at least 3 hours continuous monitoring 4.5.2. Defibrillation: at least 200 shocks at a minimum of 200J 4.5.3. Pacing: at least 2 hours continuous pacing

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 21st day of July 2022, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBA
Chairman, BAC for Pharma Supplies and Med Equipment