



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 035-2020**

**SUBJECT : Revised Technical Specification**

**PROJECT : 25,080 pcs. Disposable Isolation Gown - ITB No. 049.20**

\*\*\*\*\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above was revised as follows:

**SPECIFICATIONS**

**DISPOSABLE ISOLATION GOWN**

1. Non-woven material
2. Non-sterile
3. With AAMI (Association for the Advancement of Medical Instrumentation) certification of Level 1 or level 2 barrier performance or its equivalent
4. Dimensions:
  - a. Width – 70 cm (+/- 5cm)
  - b. Length – 110 cm (+/- 5cm)
  - c. Arm hole – 24 cm (+/- 5cm)
  - d. Sleeve length – 62 cm (+/- 5cm)
5. Opening at the back, with back fasteners (ties or other means)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 14<sup>th</sup> day of December 2020, BAC Office, PHC.

**RANULFO B. JAVELOSA, MD.**

Chairman, BAC for Infra., Med Supplies and Services