SUPPLEMENTAL BID BULLETIN No. 035-2021

SUBJECT

: Revisions in the Bidding Documents

PROJECT/S

166 units Multi - Purpose Suction Device- ITB No. 039.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally preissued bidding documents as mentioned above, revision is hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	Technical Specification 3.1. Reusable collection jar 350-500 ml, transparent	Technical Specification 3.1. Reusable collection jar 350-1,200 ml, transparent
	3.2. Provision of disposal liner for the 350 ml to 500 ml reusable collection jar	3.2. Provision of disposal liner for the 350 ml to 350 ml to 1,200 ml reusable collection jar
	3.4. Vacuum Pressure 70-360 mmHG	3.4. Vacuum Pressure 0-360 mmHG
	3.11. With suction connecting tube that is universal (compatible with any connection port)	3.11. With suction connecting tube (canister to gauge) that is universal (compatible with any connection port)
	3.15. Collection jar maybe subjected to sterilization measures for a maximum of 134°C	3.15. Collection jar maybe subjected to autoclaving
	3.16. Connecting tubes maybe subjected to sterilization measures for a maximum of 134°C	3.16. Connecting tubes maybe subjected to autoclaving
	3.19. Three (3) to five (5) sets of consumables accessories included per set	3.19 Five (5) sets of consumables accessories included per set a. Non sterile disposable liners b. connecting tube (from canister to gauge)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 9th day of July 2021, BAC Office.

RANULFOR. JAVELOSA, JR., M.D.

Chairman BAC for Infra., Med Supplies and Services

Joyette Ann F. Mindoro