



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 039-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 4 unit Defibrillator with Pacer and Cart - ITB No. 034.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	Technical Specifications-	Technical Specifications
	1.10. Output energy: 2 joules to 360; adjustable settings	1.10. Output energy: 2 joules to 270; or 2 joules - 360; adjustable settings
	1.12 Charging time of less than 5 seconds for energy of 200 joules	1.12 Charging time is 5 seconds or less for energy of 200 joules
	1.18. Built in handle that can be mounted on bed or stretcher railings	1.18. Built in handle
		1.19. Can be mounted / hooked to bed
	7. Others	7. Others: 7.2 Delivery of the items is required by 60 calendar days.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 9th day of July 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment