



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 043-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit EMG Machine - ITB No. 051.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 37	3. Technical Specifications 3.3. Handheld Stim Troller Plus with three programmable buttons, remote stim and control 3.5. USB Footswitch	3. Technical Specifications 3.3. Handheld Stim Troller Plus programmable buttons, remote stim and control and/or programmable footswitch 3.5. Footswitch (USB or other type of connection)
Page 38	3.6.e. Integrated ultrasound (MSK) 3.9. Stainless steel Cart/Trolley with handle	3.6.5. Optional Integrated ultrasound (MSK) 3.9. Metal or steel (stainless or powder coated) Cart/Trolley with handle

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 13th day of September 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment

