



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 044-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Intra Aortic Balloon Pump - ITB No. 058.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 38	B. Minimum Technical Specification 1.f.2. Size at least 12-13 inch diagonal display	1.f.2. Size at least 12 inch diagonal display
Page 39	6.a. 5-lead ECG (I, II,II,AvI, and V)	6.a. 5-lead ECG (I, II,III, aVL, aVF and V)
Page 40	3. Other Requirements 3.3. Regional/Local Standards -Philippines 3.4. Regional DOH-Philippines	- Delete - Delete

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 13th day of September 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.
Chairman, BAC for Pharma Supplies and Med Equipment

