



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 048-2023

SUBJECT : Modification in the Bidding Documents

PROJECT/S : 1 unit Hypo-Hyperthermia - ITB No. 059.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are hereby introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
	Description of Function (not part of technical specifications)	NO CHANGE
1.1.2	Dual reservoir	NO CHANGE
1.1.3	Control panel (pressure button/touch screen) with LED digital display of temperature (both patient and water) in Centigrade	Control panel (pressure button/touch screen), with provision for LED digital display of temperature (both patient and water) in Centigrade
b.	Patient Temperature Range: 30°C to 40°C	NO CHANGE
1.2.2	Provision for "repair kit"(for accidental prick)	Provision of "repair kit" (for accidental prick) OR replacement of thermal pad within the warranty period with no additional cost
3.2	Re-settable over-current breaker shall be fitted for protection	Re-settable over-current breaker shall be fitted for protection OR ITS EQUIVALENT
5.1	User/Technical Maintenance manuals to be supplied in English	User/Technical Maintenance manuals to be supplied in English (hard and soft copy) upon delivery

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 31st day of July 2023, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharmaceutical Supplies & Med Equipment