



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 049-2023

SUBJECT : Revised Bidding Documents

PROJECT : 1 unit Re-Op Instruments Set (Pedia) - ITB No. 062.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications is hereby introduced.

Attached - Revised Technical Specifications - **Annex "A"**

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 8th day of August 2023, BAC Office, PHC.


RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infrastructure, Med Supplies & Services

Item	Specifications		Statement of Compliance
	Project: 1 unit Re-Op Instruments Set (Pedia)		
A	Qty/Items	Technical Specifications	
	1 Blade Holder #3	Length= 5-5 ½ inches	
	1 Blade Holder #7	Length= 6-7 inches	
	2 DeBakey Forcep, Fine	Straight 7 ¾ – 8 inches; Tapered to a 1 mm extra fine tip	
	2 DeBakey Forcep, Regular	Straight 7 ¾ – 8 inches; Heavy jaw taper from 2.0 to 2.8 mm wide	
	1 Russian Forcep	Length= 8 inches	
	1 Nerve Hook, Adson (Blunt)	Length= 7 ¼ to 8 inches	
	1 Vein Retractor	Length= 5 ½ – 6 inches; Blade 6x11 to 12mm	
	1 Sen-Miller Retractor	Length= 6 inches; Blunt tip	
	10 Kelly Forcep	Curved; Serrated Jaw; Length= 6-6 ½ inches	
	5 Mosquito Forcep, Curved	Length= 5-5 ½ inches; Curved Jaws, Full Serration	
	5 Mosquito Forcep, Straight	Length= 5-5 ½ inches; Straight Jaws, Full Serration;	
	1 Cooley Vascular Clamp	Length= 6 to 8 inches; 52 to 55 mm Jaw length 55 to 60° angle	
	1 Cooley Vascular Clamp	Length= 6 to 7 inches; 30 to 40 mm Jaw length 60° angle	
	1 Derra C-Clamp	Length= 6 to 7 ¼ inches; 20 mm Jaw	
	1 Derra C-Clamp	Length= 6 to 7 inches; 15 to 16 mm Jaw	
	2 Mixer Forceps	Length= 7-8 inches	
	1 Schmidt Forceps	Length= 7-8 inches	
	1 Ovum Forceps, Curved	Length= 12 to 13 inches	
	2 Babcock Tissue Forcep	Length= 8-8 ½ inches	
	10 Towel clip, non-perforating	Length= 5-6 inches	
	1 Metzenbaum, regular	Length= 7 inches; with Tungsten carbide edges	
	1 Mayo Scissors, curved	Length= 5 ¾ – 6 ¾ inches; with Tungsten carbide edges	
	1 Universal Bandage Scissors	Length= 7 to 8 inches	
	1 Finochietto Rib Spreader/Retractor, Infant	Single Blade <ul style="list-style-type: none"> • Spread 54 to 145mm • Blades 12 mm deep; 15mm wide; • Arm Length = 52 to 55 mm • Length= 100 to 160 mm 	
	1 Finochietto Rib Spreader/Retractor, Junior	Single Blade <ul style="list-style-type: none"> • Spread 100 to 153 mm • Blades 28 to 45 mm deep; 15 to 38 mm wide; • Arm Length= 75 to 130 mm • Length= 100 to 170 mm 	
	1 pair (2 pcs) Army Navy Retractor	Blade Size <ul style="list-style-type: none"> • 26 x 15 mm / 45 x 15 mm • 22 x 15 mm / 39 x 15 mm 	
	1 Mayo-Hegar Needle Holder	Length= 8 inches; with Tungsten Carbide Insert	
	1 DeBakey Needle Holder	Length= 7-8 inches; with Tungsten Carbide Insert	
	1 Wire Twister	Length= 5 to 6 inches; with Tungsten Carbide Insert; Round tip	

Philippine Heart Center
East Ave., Quezon City
APPROVED

and User's Signature

	1 Wire Cutter	Length= 7 to 8 ¾ inches; Double Action Mechanism Wire cutting capacity • soft wire max. diameter 2.0 mm • hard wire max diameter 1.5 mm	
	1 Medicine cup	2 fluid ounce	
	1 Yankauer Suction Cannula	Length= 11-12 ½ inches	
	1 Andrew-Pyncho Suction Cannula	Length= 9 to 10 ½ inches	
B	Instrument Container		
	1. Bottom : Outside Dimension 470 x 274 x 90 mm (+/- 20mm on all sides)		
	1.1. Identification Label with Inscription PHC RE-OP SET PEDIA		
	1.2. Size (L x W): 56mm x 18mm		
	2. Lid: Fitted for the Bottom Container		
	2.1. With integrated germ barrier system made of PTFE (Polytetrafluoroethylene) filter valid for 5000 reprocessing cycles or Maintenance Free Barrier System		
C	Perforated Inner Basket		
	1. With fold-down handles on the side: Stainless Steel Sterilizable		
	2. Fitted insied the Bottom Instruments Container		
	3. With fitted perforated silicone mat		
D	General Specifications		
	1. Must be made of Surgical Grade Stainless Steel 300-420 Series		
	2. Instrument Etching – PHC RE-OP SET PEDIA		
	3. Warranty : 2 Years		
	4. Delivery 90 Calendar Days		

Philippine Heart Center
East Ave., Quezon City
APPROVED


End-User's Signature