



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 050-2018


SUBJECT : Modification in the Technical Specifications

PROJECT : 170 units SYRINGE PUMP

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally scheduled Project as mentioned above, modifications are hereby introduced:

Item No.	PHC Technical Specification	Clarifications/Modifications (Should read as)
Pages 74 & 75 1	Flow rate settings: 0.1 ml to 99.9 ml	Flow rate settings: 0.1 ml to 99.9 ml/hour
2	Infusion accuracy of $\pm 2-5\%$	Infusion accuracy of $\pm 1-5\%$
6	Alarm control with following alarm functions d. Occlusion detection pressure alarm e. air bubble detection f. syringe disengaged/incorrectly placed alarm g. line disconnection alarm e. nearly empty	Alarm control with following alarm functions a. Occlusion detection pressure alarm b. with or without air bubble detection c. syringe disengaged/incorrectly placed alarm d. IV tubing /line disconnection alarm e. nearly empty alarm
7	Infusion Mode Settings: b. Drip Mode	Infusion Mode Settings: b. Drip/Dose Mode
8	Plunger holder can be rotated at a minimum of 90 degrees	Syringe holder can be rotated at a minimum of 90 degrees
10	AC power source at 110-220V single phase 60Hz	AC power source at 100-220V single phase 60Hz

Approved. Done this 3rd day of October 2018. BAC Secretariat Office, PHC, East Avenue, Quezon City


CRISELLE M. GALANG, RN, MAN
Chairman, BAC for Pharma Supplies and Med Equipment