



PHILIPPINE HEART CENTER
East Avenue, Quezon City

AMENDED SUPPLEMENTAL BID BULLETIN No. 053-2023

SUBJECT : Modification in the Bidding Documents
PROJECT/S : 10 units Transport Monitor with Defibrillator
- ITB No. 065.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revision/change is hereby introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 10 & Page 41	2. Delivery Period - 60 calendar days 5.7. Delivery Period- 60 calendar days	2. Delivery Period - 90 calendar days 5.7. Delivery Period- 90 calendar days
4.5	Full charge running time 36-48 hours	Full charging to: 4.5.1. Monitoring: at least 3 hours continuous monitoring 4.5.2. Defibrillation: at least 200 shocks at a minimum of 200 joules 4.5.3. Pacing: at least 2 hours continuous pacing

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 30th day of August 2023, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharmaceutical Supplies & Med Equipment