



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 054-2021**

**SUBJECT** : Revisions in the Bidding Documents

**PROJECT/S** : 1 unit Heater Cooler Machine - ITB No. 059.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced:

<b>Item No./ Page No.</b>	<b>PHC – Technical Specifications FROM</b>	<b>Clarifications/Modifications (Should be read as) TO</b>
10	Delivery Period- 60 calendar days	Delivery Period- <b>120 calendar days</b>
22	5.3. For this purpose, contracts similar to the Project shall be: a. Heater Cooler Machine	5.3. For this purpose, contracts similar to the Project shall be: <b>a. Any other medical equipment completed contract project within the past 5 years acceptable.</b>
37	I- Technical Specification 5. 220 V, 60 Hertz electrical requirement	5. 220 V – 240 V, 60 Hertz or with a provision of frequency converter to match the hospital electrical supply.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 2<sup>nd</sup> day of November 2021, BAC Office.

**MARIETTA A. VELASCO, RN, MAN.**  
Chairman, BAC for Pharmaceutical Supplies & Med Equipment