



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 055-2019**

**SUBJECT** : Modification in the Bidding Document

**PROJECTS** : 1 Lot Cardiac Monitoring System - ITB No. 058.19

\*\*\*\*\*


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally scheduled Project as mentioned above, modifications are hereby introduced:

As Published, PhilGeps (From)	Corrected (should be read as) (To)
<b>Page 74/Item No. A.2-</b> > All units of cardiac monitors should have modular capacity and can store data.	> <b>All units of cardiac monitors should have internal storage.</b>
<b>Page 77/Item E.2:</b> > ECG reports – 3x4, 3x4 1R, 3x4 3R, 3x4 1R plus ST maps, 6x2, 12x1, Standard and Cabrera formats, plus Plan 12 Cabrera	> ECG reports – 3x4, 3x4 1R, 3x4 3R, 3x4 1R plus ST maps, 6x2, 12x1, Standard and Cabrera formats, plus Plan 12 Cabrera or its equivalent.
<b>Page 77/Item E.10:</b> > Touch screen- 1-2-3 operation/Context sensitive application/5 wire, resistive touch screen.	> Screen- touch screen or touch panel.
<b>Page 77/Item E.11:</b> > Keyboard-Back-lit 1-2-3 buttons/65 buttons standard full alphanumeric keyboard.	> Keyboard- standard full alphanumeric keyboard.
<b>Addendum</b>	> <b>Warranty of 3 years for all parts, labors and accessories.</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters/advertisement as stated therein.

Please be guided accordingly.

Approved. Done this 6<sup>th</sup> day of August 2019, BAC Office, PHC.

  
**MARIETTA A. VELASCO, RN, MAN**  
 Chairman, BAC for Pharma Supplies and Med Equipment