



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 058-2019

SUBJECT : Modification in the Bidding Document

PROJECTS : 212 units Syringe Pump - ITB No. 066.19

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally scheduled Project as mentioned above, modifications are hereby introduced:

As Published, PhilGeps (From)	Corrected (should be read as) (To)
Item No. 5 - With an LED Display	With an LED/LCD Display
Item No. 6 - Alarm Control	Audible and visual alarm function
Item No. 6b - Rapid pressure drop	Low pressure alarm (inline pressure monitoring)
Item No 7 - Drip/dose mode	Dose mode and its equivalent
Item No 11 - Equipped with built-in rechargeable lithium ion battery	Equipped with built-in rechargeable lithium or nickel metal hydride battery

This Supplemental Bid Bulletin is hereby issued to modify the original parameters/advertisement as stated therein.

Please be guided accordingly.

Approved. Done this 13th day of August 2019, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN
Chairman, BAC for Pharma Supplies and Med Equipment