



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 058-2024

SUBJECT : Modifications in the Technical Specifications of the Project

PROJECT : 1 unit Anesthesia Machine with Ventilator and
2 units High End Anesthesia Machine - ITB No. 059.24

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications are hereby introduced.

Page No/ Item No.	PHC- Technical Specification FROM	Clarifications/Amended "Should be read as" TO
Page 1 & 39	A). 1 unit Anesthesia Machine with Ventilator and B). 2 units High End Anesthesia Machine	A). 1 unit Anesthesia Machine with Ventilator and B). 2 units Fully Digital/Electronic Anesthesia Machine (High-end and latest model)
A.	1 unit Anesthesia Machine with Ventilator	
Page 40	3. Technical Specifications: 3.2. FRAME 3.2.3. With two (2) vaporizers mount compatible to Tec 7 vaporizers	3.2. FRAME 3.2.3. With two (2) vaporizers mount compatible to Tec 7 or its equivalent
	3.3. Breathing System 3.3.1. Must have a circuit volume of 2.7 L or lower. Faster emergency and most ideal breathing system for low flow anesthesia 4.3.4. 800 ml CO2 canister or lower	3.3.1. Must have a circuit volume of 2.7 L or lower or a similar system that supports 2.7 Liters volume or lower. (Faster emergency and most ideal breathing system for low flow anesthesia) 4.3.4. CO2 canister ranges from 1500 ml or lower
Page 41	3.8 Model 3.8.1. Should be latest Model 2022 up to present	3.8.1. Manufactured from 2022 and above


	<p>2. System Configuration Accessories, spare and consumable</p> <p>2.1. STORAGE/DRAWER At least two (2) drawers for accessories, dimensions at least 11.5 inch length, 13.05 inch width and 7.50 inch height</p> <p>2.2. Vaporizer One (1) unit Sevoflurane vaporizer compatible with the machine</p>	<p>2.1. STORAGE/DRAWER At least two (2) storage drawers for accessories,</p> <p>2.2. Vaporizer Two (2) units vaporizer 1` unit sevoflurane plus 1 unit isoflurane compatible with the anesthesia machine</p>
Page 42	<p>6. Documentation</p> <p>6.5. Distributorship – Exclusive</p> <p>6.7. Single Largest Completed Contract- Similar contract</p>	<p>6. Documentation</p> <p>6.5. Distributorship – Exclusive or Authorized</p> <p>6.7. Single Largest Completed Contract- Similar contract for the last 3 years</p> <p><u>Additional:</u> .6.8. Certificate of Good standing for the last 12 months</p>
Page 42	<p>B. 2 units High-End Anesthesia Machine</p> <p>1 Technical Specifications</p>	<p>B). 2 units Fully Digital/Electronic Anesthesia Machine (High-end and latest model)</p>
	<p>1.2. FRAME:</p> <p>1.2.3. With two (2) vaporizers mount compatible to Tec 7 vaporizers</p>	<p>1.2. FRAME:</p> <p>1.2.3. With two (2) vaporizers mount compatible to Tec 7 vaporizers or it equivalent</p>
Page 42	<p>1. Technical Specification:</p> <p>1.3. BREATHING SYSTEM</p> <p>1.3.1. Must have a circuit volume of 2.7 L. or lower. (Faster emergency and most ideal breathing system for low flow anesthesia)</p> <p>1.3.4. 800 ml CO2 canister or lower</p>	<p>1.3. BREATHING SYSTEM</p> <p>1.3.1. Must have a circuit volume of 2.7 L. or a similar system that supports 2.7.L volume or lower. (Faster emergency and most ideal breathing system for low flow anesthesia)</p> <p>1.3.4. CO2 canister ranges from 1500 ml or lower</p>
	<p>1.8. Model:</p> <p>1.8.1. Should be latest model 2022 up to present</p>	<p>1.8.1. Should be latest model and manufactured 2022 or up to present</p>

6. Documentation 6.5. Distributorship – Exclusive 6.7. Single Largest Completed Contract- Similar contract	6. Documentation 6.5. Distributorship – Exclusive or Authorized 6.7. Single Largest Completed Contract- Similar contract or any medical operating room equipment for the last 3 years. <u>Additional:</u> .6.8. Good standing for the last 12 months
---	--

This Supplemental Bid Bulletin is hereby supersedes the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 22nd day of July 2024, BAC Office, PHC.


ANTONIO D. PASQUAL, MD.
 Chairman, BAC for Pharmaceutical Supplies and Medical Equipment