



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 061-2021**

**SUBJECT** : Revisions in the Bidding Documents

**PROJECT/S** : **59 units Syringe Pump - ITB No. 078.21**



To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced

<b>Item No./ Page No.</b>	<b>PHC – Technical Specifications FROM</b>	<b>Clarifications/Modifications (Should be read as) TO</b>
1	1.3. Small and lightweight: approximate weight 1.2-2.1 kg	1.3. Small and lightweight: <b>approximate weight 1.2-2.5 kg</b>
2	1.10. Flow Range: 0.01-1200 ml/hr	<b>1.10 Flow Range: 0.1-1200 ml/hr</b>
3	1.19. Waterproof and/or drip proof	<b>1.19. Splash proof and/or drip proof</b>
4	1.20 Silent motor device sound of 30-40 dBA	<b>1.20 NO CHANGE-</b> This is in accordance to the recommended noise levels in hospitals by the International Noise Council and WHO <u><a href="http://www.who.int/docstore/peh/noise/Comnoise-4.pdf">http://www.who.int/docstore/peh/noise/Comnoise-4.pdf</a></u>
5	3.1 Manufacturer/Supplier should have ISO 15883 certificate to Quality Standard	3.1 Manufacturer/Supplier should have ISO <b>13485 Certificate to Quality Standard</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 19<sup>th</sup> day of November 2021, BAC Office.

  
**MARIETTA A. VELASCO, RN, MAN.**   
Chairman, BAC for Pharmaceutical Supplies & Med Equipment