



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 062-2023

SUBJECT : Modification in the Technical Specification

PROJECT/S : 1. 9 units Automated External Defibrillator - ITB No. 076-23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are hereby introduced

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended "Should be read as"
Page 39	Defibrillator: 1.10. Display Format: High resolution LCD with capacitive touch panel	Defibrillator: 1.10. Display Format: "High resolution LCD or TFT Color LCD, 800 x 480 pixels" can be considered in the technical specifications. Non- touch/Touch " With capacitive touch panel" can be deleted.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 3rd day of October 2023, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharmaceutical Supplies and Medical Equipment

