



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 063-2022

SUBJECT : Revision in the Technical Specification

PROJECT/S : 1 unit Endoscopy System - ITB.074.22

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

| Page No. | PHC – Technical Specifications FROM | Clarifications/Modifications (Should be read as) TO |
|----------|--|---|
| 37 | 1. Technical Specifications; 1.1. Video Processor (1 unit) 1.1.1. High Definition Video processor with buffer/ internal memory of 1 GB 1.1.3. With 2 HD-SDI input/output | 1.1.1. High Definition Video processor with buffer/ internal memory for recording of images and video 1.1.3. With 1 HD-SDI input/output |
| 38 | 1.3. Colonoscope Set A g. Distal end diameter - 9.2 mm h. Insertion Tube diameter - 9.3 mm | 1.3. Gastroscope Set A g. Distal end diameter - 9.0 mm -9.2 mm h. Insertion Tube diameter - 9.0 mm -9.3 mm |
| | 1.3. Colonoscope: Set B a. Field of view - 170° | 1.3. Colonoscope: Set B a. Field of view - 140° - 170° |
| | | For Inclusion: 1.5. Printer for printing of images |

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 30th day of September 2022, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment