



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 067-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 Lot Patient Call System/Call Light (Re-Bid) - ITB No. 103.21


To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 22	5.3. b. completed within the past 2 years prior to the deadline for the submission and receipt of bids.	5.3. b. completed within the past 3 years prior to the deadline for the submission and receipt of bids.
1	4.3. Suitable UPS with maintenance free batteries for minimum of one-hour back-up should be supplies with the systems.	4.3. Suitable UPS with maintenance free batteries for minimum of 30 minutes back-up should be supplies with the systems.
2	5.7. 5 years warranty in parts, accessories and services with free preventive maintenance every 6 months.	5.7. 5 years warranty in parts, accessories and services with free preventive maintenance every 6 months for 5 years .
3	6.2. Certificate of calibration and inspection	Delete
4	6.5. List of equipment available for providing calibration and routine maintenance support as per manufacturer/documentation in service/technical manual.	Delete

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 10th day of December 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN
Chairman, BAC for Pharmaceutical Supplies & Med Equipment