



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 069-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 2 units Transport Monitor (Cardiac Monitor) - ITB No. 095.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO)
Page 37	1.1. Standard parameter at least 6 leads ECG, NIBP, SPO2, RR, temp, with provision for a pressure line dedicated cables and connection	1.1. Standard parameter at least 5-6 leads ECG , NIBP, SPO2, RR, temp, with provision for a pressure line dedicated cables and connection
	1.3. Minimum of 6-12 numeric data waveform display, configuration screen display, touch screen	1.3. Minimum of 4-12 numeric data and /or waveform display, configurable screen display, touch screen
	1.8. Dual channel recorder/printer	1.8. Built-in printer
	1.9. Monitor: light weight 3-5 kgs.	1.9. Monitor: light weight 1-5 kgs.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 10th day of December 2021, BAC Office.


MARIETTA A. VELASCO, RN, MAN.

Chairman, BAC for Pharmaceutical Supplies & Med Equipment

