



PHILIPPINE HEART CENTER  
East Avenue, Quezon City

**SUPPLEMENTAL BID BULLETIN No. 073-2022**

**SUBJECT : Modifications/Amendments in the Bidding Documents**

**PROJECT/S : 2 units Infant and Pediatric High Frequency Oscillatory Ventilator (HFOV) - ITB No. 087.22**

\*\*\*\*\*

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
1.3.4	<b>Technical Specifications:</b> Standard Setting b. Respiratory Rate 1 to $\geq$ 150 bpm	Standard Setting b. Respiratory Rate <b>lower limit of 0-2 bpm</b>  <b>upper limit of 150 to 200 bpm</b>
5	<b>Standard, Safety and Training</b> 5.3. Comprehensive technical training for PHC-BIOMED (factory training) and User's training to staff for familiarity with the system	5.3. Comprehensive technical training for PHC-BIOMED and <b>End-user's for familiarity with the system by a certified training provider</b>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

**Approved.** Done this 20<sup>th</sup> day of October 2022, BAC Office, PHC.

  
**MARIETTA A. VELASCO, RN, MAN, MBAH**  
Chairman, BAC for Pharma Supplies and Medical Equipment

9