



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 073-2023

SUBJECT : Modification in the Bidding Documents

PROJECT/S : “1 Lot Supply and Delivery of Reagents, Calibrators, Controls and Consumables with Installation, Testing, Commissioning/Decommissioning and Free Use of Two (2) Units Brand New, Latest Model Fully Automated Coagulation Analyzers (Multi-Year Contract- 3 years)” - ITB No. 097.23

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents, thus, revisions/changes are hereby introduced.

Item No.	PHC- Technical Specification	Clarifications/Amended “Should be read as
	From	To
	B. Specifications for Coagulation Analyzers	B. Specifications for Coagulation Analyzers
Page 40	21. Must have an external quality assurance participation with at least very satisfactory rating	21. Must have an external quality assurance participation
	22. Must have a throughput of ≤ 3 minutes/test (PT) and ≤ 8 minutes/test (APTT)	22. Must have an average throughput of ≤ 6 minutes/test , 2 tests in throughput (PT and APTT)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 9th day of November 2023, BAC Office, PHC.


RANULFO B. JAVELOSA, JR., MD.
Chairman, BAC for Infrastructure, Medical Supplies & Services