



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 075-2021

SUBJECT : Revisions in the Bidding Documents

PROJECT/S : 1 unit Platelet Agitator with Incubator - ITB No. 099.21

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revisions are hereby introduced.

Item No./ Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO)
Page 37	1.1. Brand new, not demo unit with capacity -96 random bags/32 apheresis bags holding capacity	1.1. Brand new, not demo unit with capacity -96 random bags and 32 apheresis bags holding capacity
Page 38	1.9.5. Low battery alarm	1.9.5. No/Faulty alarm
	1.9.6. No motion alarm	1.9.6. Built-in motion alarm
	1.10. Adjustable alarm volume and tune	1.10. Adjustable alarm volume
	1.15. Power failure back-up battery of at least 2 hours	Delete
Page 39	4. Standards, Safety and Training 4.18. For Inclusion	4. Standards, Safety and Training 4.18. Technician should provide working gown during visit, performance of preventive maintenance and trouble shooting.
	5. Documentation 5.4. List of equipment available for providing calibration and routine maintenance support as per manufacturer documentation in service/technical manual.	5. Documentation 5.4. List of equipment/tools available for providing calibration and routine maintenance support as per manufacturer documentation in service/technical manual.

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 10th day of December 2021, BAC Office.

MARIETTA A. VELASCO, RN, MAN.

Chairman, BAC for Pharmaceutical Supplies & Med Equipment