



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 082-2023

SUBJECT : Revised Technical Specification

**PROJECT : 18 sets Radiation Protective Equipment Wrap around
with Thyroid Shield – ITB No. 109.23**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued bidding documents as mentioned above, revised technical specifications is hereby introduced.

Page No/ Item No.	PHC- Technical Specification	Clarifications/Amended “Should be read as“
Page 39	<p>1. Dimension and Weight:</p> <p>13. Double layer lead-free fabric (ANTIMONY AND BISMUTH)</p> <p>a. WRAP AROUND GOWN- 0.50 Pb equivalent to 0.35 mm or higher thickness of antimony and bismuth fabric.</p> <p>b. THYROID COLLAR- 0.25m Pb equivalent to 0.15 mm or higher thickness of antimony and bismuth fabric</p>	<p>1. Dimension and Weight:</p> <p>13. Double layer lead-free fabric (ANTIMONY AND BISMUTH)</p> <p>a. WRAP AROUND GOWN- 0.35 mm or higher thickness of antimony and bismuth fabric.</p> <p>b. THYROID COLLAR- 0.15 mm or higher thickness of antimony and bismuth fabric</p>

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 11th day of December 2023, BAC Office, PHC.


ANTONIO D. PASCUAL, MD.

Chairman, BAC for Pharmaceutical Supplies and Medical Equipment