



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 084 -2022

SUBJECT : Modifications/Amendments in the Bidding Documents

**PROJECT/S : 1 unit Continuous Non-Invasive Arterial Pressure Monitor
- ITB No. 107.22**

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally pre-issued technical specifications as mentioned above, revisions/changes are hereby introduced.

Page No.	PHC – Technical Specifications FROM	Clarifications/Modifications (Should be read as) TO
Page 10	Delivery Period – 30 calendar days	Delivery Period – 60 calendar days
Page 38	Technical Specifications: 1.1.10. PPV- Fluid Responsiveness: 0.1-40%	Technical Specifications: 1.1.10. PPV- Fluid Responsiveness: minimum 0.1-0.2 maximum 40%
	2. System Configuration Accessories, spares and consumables 2.2. Six (6) Non-Invasive Arterial Pressure Monitoring device (2 Small, 2 Medium, 2 Large)	2. System Configuration Accessories, spares and consumables 2.2. Six (6) Non-Invasive Arterial Pressure Monitoring finger sensor (2 Small, 2 Medium, 2 Large)

This Supplemental Bid Bulletin is hereby issued to modify the original parameters as stated therein. All other requirements previously stated on the bid documents must be complied.

Please be guided accordingly.

Approved. Done this 28th day of October 2022, BAC Office, PHC.

MARIETTA A. VELASCO, RN, MAN, MBAH
Chairman, BAC for Pharma Supplies and Medical Equipment